



AssayMax Human Complement C4 ELISA Kit

Catalog Number EC2102-1

Lot#

Introduction

Complement 4 (C4) is the second component to react in the complement sequence. It is a beta-globulin with a sedimentation coefficient of 18.7 and a molecular weight of 240,000. The C4 component participates in the initial step of activation of classical complement pathway (1). Lower C4 was associated with primary biliary cirrhosis (2), human systemic lupus erythematosus (3), and chronic liver disease (4).

Principal of the Assay

The AssayMax Human complement C4 ELISA (Enzyme-Linked Immunosorbent Assay) kit is designed for detection of human complement C4 in plasma, serum and cell culture supernatants. This assay employs a quantitative competitive sandwich enzyme immunoassay technique that measures human complement C4 in less than 2 hours. A polyclonal antibody specific for human complement C4 has been pre-coated onto a 96-well microplate with removable strips. Complement C4 in standards and samples is competed by a biotinylated complement C4 sandwiched by the immobilized antibody and streptavidin-peroxidase conjugate. All unbound material is then washed away and a peroxidase enzyme substrate is added. The color development is stopped and the intensity of the color is measured.

Caution and Warning

- This kit is for research use only.
- The kit should not be used beyond the expiration date.
- The Stop Solution is an acid solution.

Reagents

- **Human Complement C4 Microplate:** A 96-well polystyrene microplate (12 strips of 8 wells) coated with a polyclonal antibody against human complement C4.
- **Sealing Tapes:** Each kit contains 3 pre-cut, pressure-sensitive sealing tapes that can be cut to fit the format of the individual assay.
- **Human Complement C4 Standard:** Human complement C4 in a buffered protein base (40 µg, lyophilized).

- **Biotinylated Complement C4:** 1 vial, lyophilized.
- **Streptavidin-Peroxidase Conjugate (SP Conjugate):** A 100-fold concentrate (120 µl).
- **EIA Diluent Concentrate (10x):** A 10-fold concentrated buffered protein base (30 ml).
- **Wash Buffer Concentrate (10x):** A 10-fold concentrated buffered surfactant (2 x 30 ml).
- **Chromogen Substrate:** A ready-to-use stabilized peroxidase chromogen substrate tetramethylbenzidine (8 ml).
- **Stop Solution:** A 0.5 N hydroxychloric acid (12 ml) to stop the chromogen substrate reaction.

Storage Condition

- Store unopened kit at 2-8⁰C up to expiration date.
- Opened reagents may be stored for up to 1 month at 2-8⁰C. Store reconstituted standard and Biotinylated complement C4 at -20⁰C or below.
- Opened unused strip wells may return to the foil pouch with the desiccant pack, reseal along zip-seal. May be stored for up to 1 month in a vacuum desiccator.

Other Supplies Required

- Microplate reader capable of measuring absorbance at 450 nm.
- Pipettes (1-20 µl, 20-200 µl, and multiple channel).
- Deionized or distilled reagent grade water.

Sample Collection, Preparation and Storage

- **Plasma:** Collect plasma using one-tenth volume of 0.1 M sodium citrate as an anticoagulant. Centrifuge samples at 2,000x g for 10 minutes and assay. Dilute samples 1:400 into EIA Diluent. Store samples at -20⁰C or below for up to 3 months. Avoid repeated freeze-thaw cycles.
- **Serum:** Samples should be collected into a serum separator tube. After clot formation, centrifuge samples at 2,000 x g for 10 minutes. Remove serum and assay. Dilute samples 1:400 into EIA Diluent. Store serum at -20⁰C or below. Avoid repeated freeze-thaw cycles.
- **Cell Culture Supernatants:** Centrifuge cell culture media at 3,000 x g for 10 minutes to remove debris. Collect supernatants and assay. Store samples at -20⁰C or below. Avoid repeated freeze-thaw cycles.

Reagent Preparation

- Freshly dilute all reagents and bring all reagents to room temperature before use. If crystals have formed in the concentrate, mix gently until the crystals have completely dissolved.
- **Standard Curve:** Reconstitute the 40 µg of complement C4 Standard with 2 ml of EIA Diluent to generate a solution of 20 µg/ml. Allow the standard to sit for 10 minutes with gentle agitation prior to making dilutions. Prepare triplicate standard points by serially diluting the standard solution (20 µg/ml) 1:2 with EIA Diluent to produce 10, 5, 2.5, 1.25, 0.625, and 0.313 µg/ml solutions. EIA Diluent serves as the zero standard (0 µg/ml). Any remaining solution should be frozen at < -20⁰C.

Standard Point	Dilution	[Complement C4] ($\mu\text{g/ml}$)
P1	Standard (20 $\mu\text{g/ml}$)	20.000
P2	1 part P1 + 2 parts EIA Diluent	10.000
P3	1 part P2 + 2 parts EIA Diluent	5.000
P4	1 part P3 + 2 parts EIA Diluent	2.500
P5	1 part P4 + 2 parts EIA Diluent	1.250
P6	1 part P5 + 2 parts EIA Diluent	0.625
P7	1 part P6 + 2 parts EIA Diluent	0.313
P8	EIA Diluent	0.000

- **Biotinylated Complement C4:** Dilute Biotinylated complement C4 with 4 ml EIA Diluent to produce a working solution. Any remaining solution should be frozen at $< -20^{\circ}\text{C}$.
- **EIA Diluent Concentrate (10x):** Dilute the EIA Diluent 1:10 with reagent grade water.
- **Wash Buffer Concentrate (10x):** Dilute the Wash Buffer Concentrate 1:10 with reagent grade water.
- **SP Conjugate (100x):** Spin down the SP Conjugate briefly and dilute the desired amount of the conjugate 1:100 with EIA Diluent.

Assay Procedure

- Prepare all reagents, working standards and samples as instructed. Bring all reagents to room temperature before use. The assay is performed at room temperature ($20-30^{\circ}\text{C}$).
- Remove excess microplate strips from the plate frame and return them immediately to the foil pouch with desiccant inside. Reseal the pouch securely to minimize exposure to water vapor and store in a vacuum desiccator.
- Add 25 μl of standard or sample per well, and immediately add 25 μl of Biotinylated Complement C4 to each well (on top of the Standard or sample) and mix gently. Cover wells with a sealing tape and incubate for 1 hour. Start the timer after the last sample addition.
- Wash five times with 200 μl of Wash Buffer. Invert the plate and decant the contents, and hit it 4-5 times on absorbent paper towel to completely remove liquid at each step.
- Add 50 μl of Streptavidin-Peroxidase Conjugate to each well and incubate for 30 minutes. Turn on the microplate reader and set up the program in advance.
- Wash five times with 200 μl of Wash Buffer.
- Add 50 μl of Chromogen Substrate per well and incubate for about 10 minutes or till the optimal blue color density develops. Gently tap plate to ensure thorough mixing and break the bubbles in the well with pipette tip.
- Add 50 μl of Stop Solution to each well. The color will change from blue to yellow.
- Read the absorbance on a microplate reader at a wavelength of 450 nm **immediately**. Please note that after the reaction is stopped for about 10 minutes, some black particles may be generated at high concentration point, which will reduce the readings.


Data Analysis

- Calculate the mean value of the triplicate readings for each standard and sample.

- To generate a Standard Curve, plot the graph using the standard concentrations on the x-axis and the corresponding mean 450 nm absorbance on the y-axis. The best-fit line can be determined by regression analysis using log-log or 4-parameter curve fit.
- Determine the unknown sample concentration from the Standard Curve and multiply the plasma or serum value by the dilution factor of 400.

Standard Curve

- The curve is provided for illustration only. A standard curve should be generated each time the assay is performed.



Performance Characteristics

- The minimum detectable dose of complement C4 is typically 60 ng/ml.
- Intra-assay and inter-assay coefficients of variation were 4.9% and 8.3% respectively.
- No significant cross-reactivity or interference was observed.

References

1. Samano ES *et al.* (2004) *Rev Hosp Clin Fac Med Sao Paulo.* 59(3): 138-44
2. Gardinali M *et al.* (1998) *Clin Immunol Immunopathol* 87(3): 297-303
3. Yang Y *et al.* (2004) *Curr Dir Autoimmun* 7:98-132
4. Schirren CA *et al.* (1995) *Dig Dis Sci* 40(6): 1221-5

Revision 2.1.1